

# Trunk

## Safety and performance information

### Technical Specification

Dimensions	120 x 120 x 130 cm (Length x Width x Height)
Dimensions with PC Support	70 x 85 cm (Length x Height)
Weight	135 Kg
Instability levels	10
Load measurement	150 Kg
Supply	220-240 VAC, 50Hz
Absorbed power	300 W

### Functionality

- Evaluation and training in different dynamic/static orthostatic balance
- Evaluation and training of body weight transfer management in bipodal position
- Evaluation and training of cognitive and sensomotor skills of the lower limbs
- Biofeedback of the sense of position and sense of movement

### Use in orthopedic treatment (post -acute phase where feasible and/or practicable):

- Hip prosthesis
- Knee prosthesis
- Ankle prosthesis
- Ligament reconstruction (shoulders, hips, knees, ankles)
- Ligament instability and laxity (shoulders, hips, knees, ankles)
- Tendon rehabilitations of different entities
- Rachis issues
- Patellar tendon
- Degenerative problems
- Tonic recovery
- Dynamic trophism

## Trunk

### Use in neurological treatment (post-acute phase where feasible and/or practicable):

- Recovery of lower limb motricity
- Stroke
- Hemiplegia
- Ataxia
- Multiple Sclerosis (MS)
- Disorders of kinesthetic motor control
- Cerebral Palsy
- Parkinson's disease (PD)
- Paraparesis (Core Injure Plugs)
- Chronic Obstructive Pulmonary Disease (COPD)
- Degenerative problems

### Use in the norm type user (no medical use):

- Testing and training concerning posture, functional training and health fitness training;
- Testing and training for performance maintenance and enhancement;
- Test and training for the prevention and programming of adapted physical activity.

## Contraindications

### Absolute contraindications (their presence must be excluded before using the device)

- Acute myocardial infarction (in the last 2 days)
- Unstable angina pectoris
- Limited cardiac arrhythmia and/or hemodynamic
- Uncompensated / uncontrolled heart failure
- Acute pulmonary embolism or pulmonary infarction
- Myocarditis, pericarditis, acute endocarditis
- Acute aortic dissection

## Trunk

- Acute coronary syndrome
- Acute phlebothrombosis of the lower extremities
- Febrile infections
- Acute thrombosis
- Recent injuries, eg. after surgery (general surgery)
- Acute fracture
- Acute headache

**Relative contraindications (the activity can be started if the possible benefits outweigh the risk. The decision must be made by the doctor before using the device)**

- Stenosis of the left main coronary
- Main artery disease
- Heart disease of moderate severity
- Electrolyte imbalance ascertained
- Arterial hypertonia (RR> 200 mmHg syst.> 110 mmHg diast.)
- Tachyarrhythmia or bradyarrhythmia
- Hypertrophic cardiomyopathy and other forms of outflow tract obstruction
- Higher grade atrioventricular (AV) block
- Anemia
- Physical and / or mental disorders that cause inability to perform exercise properly
- Partially invasive medical devices (probes, infusers, catheters, external fixators etc.)
- Cardiac Pacemaker
- Visual impairment (vision <30% according to WHO)
- Pregnancy
- Damaged disc or traumatic spine disease
- Inflammations
- Epilepsy